
Midmark M11 - v1.0.5
Total Cycles: 1063

Name: RITTER2
Sterilizer ID: V2313228
05/03/2022 11:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.9 F 0.0
2:00 164.3 F 3.3
4:00 184.7 F 7.8
6:00 208.4 F 13.3
8:00 230.2 F 8.6
10:00 238.6 F 10.9
12:00 241.3 F 11.9
14:00 239.7 F 10.5
16:00 252.7 F 17.0
18:00 264.0 F 23.8
20:00 270.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 272.4 F 29.8
2:00 272.1 F 29.6
3:00 272.4 F 29.4
4:00 271.9 F 29.5

Min 271.5 F 29.4
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.1 F 0.0
5:00 237.3 F 0.0
10:00 224.3 F 0.0
15:00 230.7 F 0.0
20:00 224.3 F 0.0
25:00 228.0 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:30
CYCLE COMPLETE

05/03/2022 12:27 PM
ACCEPTED: X
REJECTED:
Operator: U02