
Midmark M11 - v1.0.0
Total Cycles: 5047

Name: RITTER 1
Sterilizer ID: V2178198
11/01/2024 03:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.0 F 0.0
2:00 126.9 F 0.1
4:00 146.1 F 0.5
6:00 179.2 F 2.3
8:00 206.4 F 5.9
10:00 215.9 F 10.0
12:00 225.6 F 8.7
14:00 230.5 F 8.2
16:00 241.0 F 12.8
18:00 238.5 F 9.9
20:00 237.8 F 9.0
22:00 248.4 F 14.2
24:00 259.6 F 20.3
26:00 270.2 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.6 F 28.8
2:00 272.1 F 28.3
3:00 271.3 F 28.0
4:00 272.4 F 28.8

Min 271.3 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.4 F 0.0
5:00 236.5 F 0.0
10:00 229.4 F 0.0
15:00 226.8 F 0.0
20:00 229.5 F 0.0
25:00 226.7 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 26:58
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:05:29
CYCLE COMPLETE

11/01/2024 04:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____