
Midmark M11 - v1.0.5
Total Cycles: 1272

Name: RITTER2
Sterilizer ID: V2313228
07/08/2022 08:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.8 F 0.0
2:00 116.8 F 0.5
4:00 122.9 F 2.6
6:00 147.1 F 6.0
8:00 172.3 F 10.2
10:00 214.5 F 5.0
12:00 216.9 F 7.1
14:00 234.4 F 11.4
16:00 230.4 F 7.5
18:00 240.6 F 12.8
20:00 240.0 F 11.0
22:00 252.1 F 17.0
24:00 262.7 F 23.3
26:00 272.7 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.6
1:00 271.8 F 29.7
2:00 271.7 F 29.9
3:00 271.8 F 29.8
4:00 271.4 F 29.6

Min 271.4 F 29.5
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.0 F 0.0
5:00 227.6 F 0.0
10:00 222.1 F 0.0
15:00 223.5 F 0.0
20:00 221.2 F 0.0
25:00 232.6 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 26:20
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:04:41
CYCLE COMPLETE

07/08/2022 10:05 AM
ACCEPTED: X
REJECTED:
Operator: U02