

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1113

Name: RITTER 1  
Sterilizer ID: V2178198  
08/24/2021 08:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.3 F 0.0  
2:00 126.9 F 0.0  
4:00 150.9 F 0.0  
6:00 182.8 F 0.0  
8:00 206.8 F 0.4  
10:00 217.7 F 2.2  
12:00 226.1 F 4.8  
14:00 234.7 F 8.2  
16:00 244.3 F 12.5  
18:00 239.4 F 9.9  
20:00 237.5 F 8.8  
22:00 242.3 F 10.8  
24:00 248.3 F 14.1  
26:00 260.8 F 21.0  
28:00 271.7 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 27.9  
1:00 272.5 F 28.7  
2:00 272.1 F 28.1  
3:00 272.2 F 28.4  
4:00 272.7 F 28.5

Min 271.5 F 27.8  
Max 272.9 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 172.6 F 0.0  
5:00 224.7 F 0.0  
10:00 191.7 F 0.0  
15:00 194.4 F 0.0  
20:00 193.0 F 0.0  
25:00 186.8 F 0.0  
30:00 187.2 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 28:28  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 01:06:60  
CYCLE COMPLETE

08/24/2021 09:49 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02