
Midmark M11 - v1.0.0
Total Cycles: 3948

Name: RITTER 1
Sterilizer ID: V2178198
10/27/2023 07:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.0 F 0.0
2:00 69.7 F 0.2
4:00 80.1 F 1.1
6:00 110.0 F 3.2
8:00 137.0 F 6.0
10:00 165.7 F 9.6
12:00 194.6 F 10.9
14:00 207.6 F 4.4
16:00 218.9 F 7.7
18:00 231.2 F 11.5
20:00 229.2 F 6.7
22:00 241.8 F 11.9
24:00 237.7 F 9.4
26:00 250.8 F 15.5
28:00 262.0 F 22.0
30:00 272.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.5 F 28.8
2:00 271.5 F 28.5
3:00 272.2 F 28.7
4:00 271.5 F 28.5

Min 271.1 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 196.8 F 0.0
10:00 220.4 F 0.0
15:00 219.3 F 0.0
20:00 220.2 F 0.0
25:00 224.1 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:28
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:08:40
CYCLE COMPLETE

10/27/2023 08:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____