

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4203

Name: RITTER2  
Sterilizer ID: V2313228  
12/11/2024 10:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.5 F 0.0  
2:00 178.4 F 2.1  
4:00 181.5 F 6.7  
6:00 205.7 F 12.1  
8:00 222.9 F 6.7  
10:00 236.4 F 12.0  
12:00 236.1 F 9.8  
14:00 237.3 F 9.2  
16:00 249.5 F 15.4  
18:00 261.5 F 22.2  
20:00 272.4 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.9  
1:00 272.4 F 29.3  
2:00 271.9 F 29.3  
3:00 272.3 F 29.2  
4:00 272.0 F 29.4

Min 271.6 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.4 F 0.0  
5:00 243.1 F 0.0  
10:00 226.1 F 0.0  
15:00 224.5 F 0.0  
20:00 223.6 F 0.0  
25:00 221.4 F 0.0  
30:00 213.3 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 20:26  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 00:59:44  
CYCLE COMPLETE

12/11/2024 11:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_