

-----  
Midmark M11 - v1.0.5  
Total Cycles: 609

Name: RITTER2  
Sterilizer ID: V2313228  
12/08/2021 07:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.2 F 0.0  
2:00 75.2 F 0.0  
4:00 75.6 F 0.1  
6:00 92.6 F 0.3  
8:00 140.4 F 3.2  
10:00 181.6 F 6.8  
12:00 185.3 F 10.9  
14:00 209.0 F 4.8  
16:00 221.7 F 8.0  
18:00 231.2 F 12.0  
20:00 232.4 F 8.5  
22:00 240.3 F 11.9  
24:00 243.4 F 12.4  
26:00 255.2 F 18.5  
28:00 264.8 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 271.8 F 29.6  
2:00 272.0 F 30.0  
3:00 271.8 F 29.7  
4:00 271.8 F 29.8

Min 270.5 F 28.8  
Max 273.4 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.6 F 0.0  
5:00 240.2 F 0.0  
10:00 220.3 F 0.0  
15:00 221.2 F 0.0  
20:00 220.6 F 0.0  
25:00 227.0 F 0.0  
30:00 220.1 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 29:39  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 01:07:26  
CYCLE COMPLETE

12/08/2021 08:45 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02