

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3484

Name: RITTER 1  
Sterilizer ID: V2178198  
07/05/2023 07:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.9 F 0.0  
2:00 71.1 F 0.0  
4:00 84.0 F 0.0  
6:00 109.2 F 1.1  
8:00 138.6 F 2.3  
10:00 166.4 F 4.2  
12:00 181.4 F 6.9  
14:00 197.3 F 10.5  
16:00 214.3 F 7.0  
18:00 220.9 F 7.3  
20:00 231.2 F 11.2  
22:00 228.5 F 6.4  
24:00 241.0 F 11.9  
26:00 237.7 F 9.5  
28:00 250.5 F 15.6  
30:00 261.2 F 21.7  
32:00 271.3 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.6 F 29.2  
2:00 272.0 F 28.5  
3:00 271.5 F 28.6  
4:00 272.6 F 29.1  
  
Min 271.2 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.4 F 0.0  
5:00 209.5 F 0.0  
10:00 219.8 F 0.0  
15:00 233.8 F 0.0  
20:00 213.9 F 0.0  
25:00 212.9 F 0.0  
30:00 211.5 F 0.0  
  
Drying Complete

FILLING: 0:46  
HEATING: 32:44  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:10:43  
CYCLE COMPLETE

07/05/2023 08:40 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_