
Midmark M11 - v1.0.5
Total Cycles: 1119

Name: RITTER2
Sterilizer ID: V2313228
05/26/2022 07:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.5 F 0.0
2:00 193.5 F 2.5
4:00 197.3 F 7.4
6:00 221.7 F 13.5
8:00 228.9 F 9.6
10:00 234.7 F 8.9
12:00 242.1 F 11.9
14:00 245.4 F 13.5
16:00 258.3 F 20.6
18:00 269.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.6 F 29.6
2:00 272.0 F 29.6
3:00 272.5 F 29.5
4:00 272.0 F 29.5

Min 271.7 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 199.9 F 0.0
10:00 202.7 F 0.0
15:00 199.4 F 0.0
20:00 192.0 F 0.0
25:00 189.5 F 0.0
30:00 184.5 F 0.0
Drying Complete

FILLING: 2:40
HEATING: 18:57
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

05/26/2022 08:50 AM
ACCEPTED: X
REJECTED:
Operator: U02