
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001794
7:29 19/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:53
STERILIZING 3:59
136.5 C 222kPa 3:59
136.4 C 224kPa 4:29
136.4 C 222kPa 4:59
136.3 C 223kPa 5:29
136.1 C 221kPa 5:59
136.1 C 222kPa 6:29
136.1 C 221kPa 6:59
136.4 C 224kPa 7:29
136.4 C 224kPa 7:29
VENTING 7:29
AIR DRYING 8:19
CYCLE COMPLETE 22:52

Digital Signature #
1F9015744330251D
