
Midmark M11 - v1.0.0
Total Cycles: 2273

Name: RITTER 1
Sterilizer ID: V2178198
08/31/2022 12:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.5 F 0.0
2:00 201.6 F 1.0
4:00 210.4 F 5.8
6:00 222.0 F 11.2
8:00 227.3 F 6.9
10:00 240.6 F 12.6
12:00 239.5 F 10.6
14:00 237.3 F 9.2
16:00 251.6 F 16.2
18:00 263.4 F 23.0
20:00 272.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.8 F 29.0
2:00 271.7 F 28.3
3:00 272.9 F 29.2
4:00 272.2 F 28.6

Min 271.5 F 28.2
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 232.7 F 0.0
10:00 226.6 F 0.0
15:00 226.0 F 0.0
20:00 228.5 F 0.0
25:00 229.8 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 20:12
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:60
CYCLE COMPLETE

08/31/2022 01:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____