
Midmark M11 - v1.0.0
Total Cycles: 1591

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2022 03:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.1 F 0.0
2:00 183.4 F 3.7
4:00 195.6 F 8.7
6:00 217.5 F 7.1
8:00 227.5 F 9.3
10:00 234.2 F 9.0
12:00 242.8 F 12.4
14:00 242.0 F 11.2
16:00 255.6 F 18.1
18:00 267.3 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.8
2:00 271.5 F 28.2
3:00 272.7 F 29.0
4:00 271.9 F 28.4

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 228.9 F 0.0
10:00 224.6 F 0.0
15:00 223.4 F 0.0
20:00 223.2 F 0.0
25:00 224.0 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:33
CYCLE COMPLETE

02/02/2022 04:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____