
Midmark M11 - v1.0.0
Total Cycles: 2869

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2023 01:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.5 F 0.0
2:00 134.6 F 2.2
4:00 157.2 F 5.3
6:00 179.7 F 9.5
8:00 202.3 F 7.2
10:00 218.1 F 7.1
12:00 231.1 F 11.8
14:00 232.1 F 8.2
16:00 242.5 F 12.0
18:00 243.5 F 12.3
20:00 256.1 F 18.9
22:00 267.3 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 29.5
2:00 271.8 F 28.7
3:00 272.0 F 29.3
4:00 272.4 F 29.3

Min 271.4 F 28.5
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 222.7 F 0.0
10:00 223.8 F 0.0
15:00 225.1 F 0.0
20:00 229.0 F 0.0
25:00 228.8 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 23:25
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:36
CYCLE COMPLETE

02/07/2023 02:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____