

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3222

Name: RITTER2  
Sterilizer ID: V2313228  
12/15/2023 10:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.2 F 0.0  
2:00 162.2 F 0.8  
4:00 165.6 F 3.6  
6:00 187.2 F 7.4  
8:00 205.8 F 11.5  
10:00 220.3 F 5.0  
12:00 228.9 F 9.4  
14:00 235.3 F 10.3  
16:00 240.9 F 11.6  
18:00 239.2 F 10.2  
20:00 251.8 F 16.4  
22:00 262.7 F 22.8  
24:00 272.6 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.6 F 29.4  
2:00 271.7 F 28.7  
3:00 272.0 F 29.4  
4:00 272.6 F 29.2

Min 271.2 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.4 F 0.0  
5:00 226.6 F 0.0  
10:00 213.5 F 0.0  
15:00 205.9 F 0.0  
20:00 200.6 F 0.0  
25:00 195.5 F 0.0  
30:00 201.4 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 24:27  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:03:17  
CYCLE COMPLETE

12/15/2023 11:34 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_