

-----  
Midmark M11 - v1.0.5  
Total Cycles: 623

Name: RITTER2  
Sterilizer ID: V2313228  
12/09/2021 03:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.7 F 0.0  
2:00 178.3 F 3.9  
4:00 194.9 F 9.0  
6:00 214.2 F 8.0  
8:00 228.3 F 9.6  
10:00 234.1 F 9.0  
12:00 244.1 F 13.5  
14:00 244.5 F 13.0  
16:00 257.8 F 20.0  
18:00 269.3 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.4 F 29.3  
2:00 272.4 F 29.8  
3:00 272.0 F 29.1  
4:00 272.5 F 29.8  
  
Min 271.6 F 28.9  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.6 F 0.0  
5:00 216.4 F 0.0  
10:00 223.2 F 0.0  
15:00 225.5 F 0.0  
20:00 224.4 F 0.0  
25:00 221.2 F 0.0  
Dry Cycle Aborted

FILLING: 2:28  
HEATING: 18:58  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 27:16  
TOTAL CYCLE: 00:55:44

12/09/2021 04:18 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02