
Midmark M11 - v1.0.5
Total Cycles: 3289

Name: RITTER2
Sterilizer ID: V2313228
01/11/2024 12:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.2 F 0.0
2:00 153.3 F 1.7
4:00 173.0 F 4.9
6:00 191.9 F 8.9
8:00 209.4 F 12.7
10:00 219.7 F 8.0
12:00 232.5 F 13.0
14:00 238.1 F 10.4
16:00 235.5 F 8.6
18:00 249.0 F 15.3
20:00 260.3 F 21.8
22:00 270.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.5
1:00 272.7 F 29.9
2:00 271.5 F 29.7
3:00 272.5 F 29.9
4:00 271.2 F 29.6

Min 271.2 F 29.4
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 165.1 F 0.0
5:00 245.0 F 0.0
10:00 210.3 F 0.0
15:00 192.6 F 0.0
20:00 188.1 F 0.0
25:00 185.7 F 0.0
30:00 184.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 23:34
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:02:16
CYCLE COMPLETE

01/11/2024 01:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____