

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3245

Name: RITTER 1  
Sterilizer ID: V2178198  
05/16/2023 03:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.4 F 0.0  
2:00 195.3 F 4.2  
4:00 208.9 F 9.5  
6:00 221.7 F 6.4  
8:00 234.0 F 10.8  
10:00 234.7 F 8.5  
12:00 239.3 F 10.0  
14:00 248.3 F 14.5  
16:00 261.1 F 21.7  
18:00 272.6 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.4 F 29.0  
2:00 271.7 F 28.7  
3:00 272.4 F 29.0  
4:00 271.5 F 28.4  
  
Min 271.5 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.5 F 0.0  
5:00 221.5 F 0.0  
10:00 231.1 F 0.0  
15:00 229.7 F 0.0  
20:00 228.1 F 0.0  
25:00 228.5 F 0.0  
30:00 228.3 F 0.0  
Drying Complete

FILLING: 2:20  
HEATING: 18:24  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:58:03  
CYCLE COMPLETE

05/16/2023 04:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_