
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001616
9:30 11/25/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:28
PRESSURIZING 6:22
STERILIZING 7:56
136.0 C 220kPa 7:56
136.5 C 223kPa 8:26
136.2 C 221kPa 8:56
136.3 C 222kPa 9:26
136.3 C 221kPa 9:57
136.4 C 222kPa 10:27
136.5 C 223kPa 10:57
136.3 C 221kPa 11:26
VENTING 11:26
AIR DRYING 12:48
CYCLE COMPLETE 20:34

Digital Signature #
0F797EC51462EABF
