
Midmark M11 - v1.0.0
Total Cycles: 5291

Name: RITTER 1
Sterilizer ID: V2178198
01/29/2025 11:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.4 F 0.0
2:00 176.4 F 2.6
4:00 179.0 F 6.2
6:00 198.2 F 10.1
8:00 221.7 F 6.8
10:00 220.6 F 6.9
12:00 232.5 F 11.2
14:00 231.1 F 7.3
16:00 243.8 F 12.9
18:00 241.5 F 11.1
20:00 253.3 F 17.1
22:00 263.9 F 23.4
24:00 271.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.7 F 29.7
2:00 272.7 F 29.4
3:00 271.8 F 28.9
4:00 271.6 F 29.1

Min 271.3 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 234.8 F 0.0
10:00 197.8 F 0.0
15:00 191.4 F 0.0
20:00 244.3 F 0.0
25:00 231.0 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 25:04
STERILIZING: 4:00
VENTING: 3:60
DRYING: 30:00
TOTAL CYCLE: 01:04:10
CYCLE COMPLETE

01/29/2025 12:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____