

-----  
Midmark M11 - v1.0.5  
Total Cycles: 853

Name: RITTER2  
Sterilizer ID: V2313228  
03/01/2022 02:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.4 F 0.0  
2:00 145.1 F 2.2  
4:00 162.7 F 5.6  
6:00 190.6 F 10.1  
8:00 217.3 F 5.2  
10:00 230.9 F 9.6  
12:00 235.1 F 9.4  
14:00 241.3 F 12.1  
16:00 240.1 F 10.8  
18:00 253.7 F 17.5  
20:00 265.4 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.3 F 29.3  
2:00 271.5 F 29.3  
3:00 272.4 F 29.2  
4:00 271.3 F 29.1

Min 271.1 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.0 F 0.0  
5:00 229.3 F 0.0  
10:00 226.4 F 0.0  
15:00 233.3 F 0.0  
20:00 232.4 F 0.0  
25:00 225.0 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 1:24  
HEATING: 21:43  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:18  
CYCLE COMPLETE

03/01/2022 04:06 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02