

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1674

Name: RITTER2  
Sterilizer ID: V2313228  
11/02/2022 06:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.6 F 0.0  
2:00 189.0 F 3.9  
4:00 202.3 F 9.7  
6:00 219.5 F 5.5  
8:00 235.7 F 11.7  
10:00 237.6 F 9.9  
12:00 236.9 F 9.0  
14:00 252.8 F 16.7  
16:00 265.4 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.5  
1:00 271.8 F 28.5  
2:00 271.9 F 28.4  
3:00 272.4 F 28.7  
4:00 272.8 F 29.0  
  
Min 271.7 F 28.4  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.6 F 0.0  
5:00 214.3 F 0.0  
10:00 234.4 F 0.0  
15:00 232.2 F 0.0  
20:00 221.2 F 0.0  
25:00 214.3 F 0.0  
30:00 211.3 F 0.0  
Drying Complete

FILLING: 2:25  
HEATING: 17:33  
STERILIZING: 4:00  
VENTING: 2:57  
DRYING: 30:00  
TOTAL CYCLE: 00:56:55  
CYCLE COMPLETE

11/03/2022 06:43 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02