
Midmark M11 - v1.0.5
Total Cycles: 2864

Name: RITTER2
Sterilizer ID: V2313228
09/11/2023 04:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.3 F 0.0
2:00 171.6 F 3.2
4:00 192.7 F 7.4
6:00 208.3 F 12.2
8:00 222.7 F 7.0
10:00 233.8 F 12.1
12:00 235.7 F 9.7
14:00 238.0 F 9.5
16:00 248.7 F 15.1
18:00 260.2 F 21.5
20:00 270.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.4 F 29.5
2:00 271.6 F 29.3
3:00 272.3 F 29.4
4:00 271.5 F 29.1

Min 271.4 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 213.9 F 0.0
10:00 213.1 F 0.0
15:00 209.3 F 0.0
20:00 205.5 F 0.0
25:00 205.4 F 0.0
30:00 203.8 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 20:45
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:00:32
CYCLE COMPLETE

09/11/2023 05:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____