

-----  
Midmark M11 - v1.0.5  
Total Cycles: 38

Name: RITTER2  
Sterilizer ID: V2313228  
05/20/2021 11:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.4 F 0.0  
2:00 159.1 F 2.8  
4:00 177.6 F 6.3  
6:00 198.7 F 10.6  
8:00 216.4 F 5.3  
10:00 226.2 F 9.4  
12:00 235.1 F 10.3  
14:00 240.4 F 12.1  
16:00 240.8 F 11.1  
18:00 252.8 F 17.2  
20:00 263.5 F 23.6  
22:00 272.7 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.5 F 29.8  
2:00 272.4 F 29.3  
3:00 271.5 F 28.9  
4:00 272.1 F 29.7

Min 271.5 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.3 F 0.0  
5:00 211.4 F 0.0  
10:00 224.4 F 0.0  
15:00 222.9 F 0.0  
20:00 213.3 F 0.0  
25:00 211.1 F 0.0  
30:00 208.3 F 0.0  
Drying Complete

FILLING: 1:36  
HEATING: 22:18  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:01:19  
CYCLE COMPLETE

05/20/2021 12:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_