

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4489

Name: RITTER2  
Sterilizer ID: V2313228  
04/04/2025 02:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.2 F 0.0  
2:00 122.9 F 1.6  
4:00 145.4 F 4.7  
6:00 174.0 F 9.0  
8:00 198.5 F 10.5  
10:00 217.1 F 5.9  
12:00 231.3 F 10.4  
14:00 229.8 F 7.6  
16:00 242.0 F 12.4  
18:00 240.6 F 10.7  
20:00 253.1 F 17.1  
22:00 264.4 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.2 F 28.9  
2:00 272.1 F 29.4  
3:00 271.4 F 28.7  
4:00 272.4 F 29.0

Min 271.3 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.4 F 0.0  
5:00 228.6 F 0.0  
10:00 200.4 F 0.0  
15:00 198.3 F 0.0  
20:00 192.7 F 0.0  
25:00 189.7 F 0.0  
30:00 187.9 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 23:53  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:02:23  
CYCLE COMPLETE

04/04/2025 03:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_