
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001641
13:34 30/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:19
PRESSURIZING 2:26
STERILIZING 3:30
136.0 C 220kPa 3:30
136.1 C 222kPa 4:00
136.3 C 224kPa 4:30
136.2 C 221kPa 5:00
136.4 C 225kPa 5:30
136.3 C 221kPa 6:00
136.4 C 221kPa 6:30
136.4 C 223kPa 7:00
136.4 C 223kPa 7:00
VENTING 7:00
AIR DRYING 7:50
CYCLE COMPLETE 8:08

Digital Signature #
A8BF835E8D70F935
