

-----  
Midmark M11 - v1.0.5  
Total Cycles: 389

Name: RITTER2  
Sterilizer ID: V2313228  
09/23/2021 10:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
2:00 178.6 F 3.7  
4:00 192.2 F 8.5  
6:00 212.3 F 9.3  
8:00 224.7 F 9.4  
10:00 233.8 F 9.5  
12:00 241.4 F 12.9  
14:00 241.9 F 12.0  
16:00 255.1 F 18.8  
18:00 266.6 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.3  
1:00 272.0 F 29.8  
2:00 272.2 F 29.7  
3:00 272.1 F 29.5  
4:00 272.1 F 29.4  
  
Min 271.6 F 29.2  
Max 272.5 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.1 F 0.0  
5:00 237.3 F 0.0  
10:00 226.9 F 0.0  
15:00 222.8 F 0.0  
20:00 217.0 F 0.0  
25:00 213.2 F 0.0  
30:00 214.7 F 0.0  
Drying Complete

FILLING: 2:10  
HEATING: 19:21  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:42  
CYCLE COMPLETE

09/23/2021 11:08 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02