

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3493

Name: RITTER 1  
Sterilizer ID: V2178198  
07/06/2023 08:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.4 F 0.0  
2:00 181.3 F 0.0  
4:00 193.8 F 1.2  
6:00 210.1 F 5.5  
8:00 227.9 F 10.3  
10:00 227.1 F 6.4  
12:00 235.5 F 10.9  
14:00 232.7 F 7.9  
16:00 242.6 F 11.3  
18:00 245.7 F 13.1  
20:00 258.2 F 19.7  
22:00 269.2 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.4 F 28.8  
2:00 271.3 F 28.2  
3:00 272.4 F 29.0  
4:00 271.6 F 28.3

Min 271.2 F 28.1  
Max 272.4 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
5:00 210.7 F 0.0  
10:00 233.6 F 0.0  
15:00 230.3 F 0.0  
20:00 227.8 F 0.0  
25:00 233.3 F 0.0  
30:00 225.2 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 23:04  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:01:04  
CYCLE COMPLETE

07/06/2023 09:56 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_