
Midmark M11 - v1.0.5
Total Cycles: 1756

Name: RITTER2
Sterilizer ID: V2313228
11/22/2022 07:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.5 F 0.0
2:00 81.6 F 0.0
4:00 83.2 F 0.0
6:00 113.8 F 0.0
8:00 160.5 F 0.1
10:00 194.2 F 0.5
12:00 211.7 F 4.0
14:00 225.2 F 7.5
16:00 234.1 F 11.3
18:00 231.9 F 7.0
20:00 243.3 F 12.6
22:00 240.6 F 10.9
24:00 238.5 F 9.8
26:00 252.5 F 16.7
28:00 264.1 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 272.3 F 29.3
2:00 272.1 F 28.6
3:00 272.0 F 29.2
4:00 272.5 F 28.9

Min 271.2 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.5 F 0.0
5:00 190.1 F 0.0
10:00 204.5 F 0.0
15:00 200.3 F 0.0
20:00 202.1 F 0.0
25:00 202.9 F 0.0
30:00 206.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 29:51
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:07:54
CYCLE COMPLETE

11/22/2022 08:50 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____