

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4676

Name: RITTER 1  
Sterilizer ID: V2178198  
06/26/2024 08:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.4 F 0.0  
2:00 75.0 F 0.3  
4:00 85.3 F 1.3  
6:00 116.3 F 3.5  
8:00 149.2 F 6.2  
10:00 178.9 F 9.3  
12:00 190.3 F 13.1  
14:00 205.1 F 3.5  
16:00 216.1 F 6.7  
18:00 225.7 F 10.2  
20:00 231.7 F 9.1  
22:00 237.3 F 10.4  
24:00 235.8 F 8.3  
26:00 247.3 F 14.0  
28:00 258.7 F 20.2  
30:00 269.1 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.0 F 28.4  
2:00 272.6 F 28.9  
3:00 272.0 F 28.8  
4:00 271.7 F 28.3

Min 271.5 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.7 F 0.0  
5:00 195.9 F 0.0  
10:00 205.0 F 0.0  
15:00 203.9 F 0.0  
20:00 214.3 F 0.0  
25:00 216.5 F 0.0  
30:00 214.7 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 31:02  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:09:11  
CYCLE COMPLETE

06/26/2024 09:27 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_