

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2370

Name: RITTER2  
Sterilizer ID: V2313228  
05/15/2023 04:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.3 F 0.0  
2:00 192.0 F 3.8  
4:00 211.6 F 9.4  
6:00 227.7 F 8.4  
8:00 231.1 F 11.3  
10:00 233.1 F 9.0  
12:00 239.0 F 9.9  
14:00 247.8 F 14.6  
16:00 259.9 F 21.2  
18:00 270.6 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.3  
1:00 271.9 F 29.2  
2:00 271.9 F 29.2  
3:00 272.3 F 29.3  
4:00 272.5 F 29.5  
  
Min 271.6 F 29.2  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.0 F 0.0  
5:00 218.6 F 0.0  
10:00 236.7 F 0.0  
15:00 233.0 F 0.0  
20:00 228.1 F 0.0  
25:00 229.8 F 0.0  
30:00 229.5 F 0.0  
  
Drying Complete

FILLING: 2:31  
HEATING: 19:24  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 00:59:19  
CYCLE COMPLETE

05/15/2023 05:06 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_