
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002044
7:59 15/07/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:19
PRESSURIZING 1:25
STERILIZING 2:23
136.4 C 221kPa 2:23
136.2 C 222kPa 2:53
136.4 C 222kPa 3:23
136.1 C 218kPa 3:53
136.4 C 224kPa 4:23
136.1 C 217kPa 4:53
136.4 C 223kPa 5:23
136.1 C 221kPa 5:53
136.1 C 221kPa 5:53
VENTING 5:53
AIR DRYING 6:47
CYCLE COMPLETE 6:56

Digital Signature #
B583D34ADD259E69
