
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000778
7:54 29/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:25
STERILIZING 3:31
136.0 C 228kPa 3:31
136.1 C 231kPa 4:01
136.2 C 231kPa 4:31
136.2 C 230kPa 5:01
136.3 C 233kPa 5:31
136.4 C 233kPa 6:01
136.3 C 232kPa 6:31
136.3 C 232kPa 7:01
136.3 C 232kPa 7:01
VENTING 7:01
AIR DRYING 7:52
CYCLE COMPLETE 7:54

Digital Signature #
2D9E06F3713C1782
