
Midmark M11 - v1.0.0
Total Cycles: 4718

Name: RITTER 1
Sterilizer ID: V2178198
07/11/2024 02:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.3 F 0.0
2:00 180.2 F 3.0
4:00 204.8 F 7.3
6:00 216.3 F 12.1
8:00 220.8 F 5.7
10:00 231.6 F 10.2
12:00 237.0 F 8.3
14:00 242.4 F 12.5
16:00 241.2 F 11.1
18:00 253.9 F 17.4
20:00 264.8 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.7 F 29.2
2:00 271.9 F 28.5
3:00 271.5 F 28.7
4:00 272.7 F 29.2

Min 271.2 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 218.2 F 0.0
10:00 233.0 F 0.0
15:00 229.8 F 0.0
20:00 226.8 F 0.0
25:00 222.7 F 0.0
30:00 218.4 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 21:57
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:01:24
CYCLE COMPLETE

07/11/2024 03:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____