
Midmark M11 - v1.0.0
Total Cycles: 4987

Name: RITTER 1
Sterilizer ID: V2178198
10/10/2024 02:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.3 F 0.0
2:00 143.0 F 0.4
4:00 155.2 F 1.6
6:00 183.9 F 5.6
8:00 204.3 F 10.2
10:00 223.3 F 6.0
12:00 227.7 F 9.7
14:00 235.0 F 8.6
16:00 242.0 F 12.6
18:00 241.1 F 11.3
20:00 254.3 F 17.9
22:00 265.9 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.2 F 29.2
2:00 271.3 F 28.4
3:00 272.2 F 29.3
4:00 271.7 F 28.7

Min 271.2 F 28.4
Max 272.2 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 211.2 F 0.0
10:00 210.6 F 0.0
15:00 224.5 F 0.0
20:00 221.5 F 0.0
25:00 225.5 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 23:40
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:02:11
CYCLE COMPLETE

10/10/2024 03:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____