

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000937  
12:09 17/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:57  
PRESSURIZING 4:00  
STERILIZING 5:36  
136.0 C 227kPa 5:36  
135.9 C 230kPa 6:06  
136.2 C 230kPa 6:36  
136.1 C 231kPa 7:06  
136.4 C 233kPa 7:36  
136.2 C 230kPa 8:06  
136.5 C 232kPa 8:36  
136.3 C 231kPa 9:06  
136.3 C 231kPa 9:06  
VENTING 9:06  
AIR DRYING 10:05  
CYCLE COMPLETE 21:24  
-----

Digital Signature #  
CDFFE733AE404EC3  
-----