

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002442  
11:44 23/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:36  
STERILIZING 3:28  
136.0 C 221kPa 3:28  
136.4 C 221kPa 3:58  
136.1 C 224kPa 4:28  
136.4 C 225kPa 4:58  
136.4 C 225kPa 5:28  
136.4 C 226kPa 5:58  
136.4 C 226kPa 6:28  
136.3 C 224kPa 6:58  
136.3 C 224kPa 6:58  
VENTING 6:58  
AIR DRYING 7:48  
CYCLE COMPLETE 34:41  
-----

Digital Signature #  
8AD6AB7FBD8A7AD9  
-----