
Midmark M11 - v1.0.0
Total Cycles: 3252

Name: RITTER 1
Sterilizer ID: V2178198
05/17/2023 02:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	130.9 F	0.0
2:00	135.8 F	0.0
4:00	164.0 F	0.3
6:00	190.6 F	4.4
8:00	203.7 F	9.3
10:00	221.0 F	9.4
12:00	229.3 F	8.6
14:00	239.8 F	12.3
16:00	240.4 F	11.2
18:00	238.2 F	9.6
20:00	252.3 F	16.4
22:00	264.0 F	23.3
24:00	271.4 F	28.2

mm:ss	Degrees	PSI
0:00	271.4 F	28.3
1:00	272.6 F	29.0
2:00	271.8 F	28.2
3:00	272.1 F	28.9
4:00	272.4 F	28.8

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

mm:ss	Degrees	PSI
0:00	200.2 F	0.0
5:00	228.7 F	0.0
10:00	205.4 F	0.0
15:00	204.0 F	0.0
20:00	212.5 F	0.0
25:00	233.7 F	0.0
30:00	225.8 F	0.0

Drying Complete

FILLING: 0:53
HEATING: 24:02
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:12
CYCLE COMPLETE

05/17/2023 03:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____