

-----  
Midmark M11 - v1.0.5  
Total Cycles: 711

Name: RITTER2  
Sterilizer ID: V2313228  
01/20/2022 09:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.6 F 0.0  
2:00 163.2 F 2.9  
4:00 178.4 F 7.3  
6:00 202.4 F 12.8  
8:00 218.0 F 7.5  
10:00 233.9 F 13.5  
12:00 239.1 F 11.3  
14:00 239.4 F 10.4  
16:00 253.8 F 17.7  
18:00 266.0 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.5 F 29.7  
2:00 272.1 F 29.0  
3:00 271.3 F 28.8  
4:00 272.4 F 29.7  
  
Min 271.2 F 28.6  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.1 F 0.0  
5:00 204.8 F 0.0  
10:00 208.4 F 0.0  
15:00 206.6 F 0.0  
20:00 209.8 F 0.0  
25:00 217.7 F 0.0  
30:00 212.0 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 19:39  
STERILIZING: 4:00  
VENTING: 2:57  
DRYING: 30:00  
TOTAL CYCLE: 00:57:58  
CYCLE COMPLETE

01/20/2022 09:57 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_