
Midmark M11 - v1.0.5
Total Cycles: 3684

Name: RITTER2
Sterilizer ID: V2313228
06/11/2024 08:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 70.3 F 0.1
4:00 75.8 F 0.6
6:00 98.1 F 2.3
8:00 133.7 F 4.7
10:00 161.6 F 7.7
12:00 192.2 F 11.1
14:00 201.1 F 3.1
16:00 214.4 F 6.6
18:00 223.9 F 10.4
20:00 229.9 F 8.5
22:00 237.3 F 11.2
24:00 235.5 F 9.0
26:00 248.7 F 15.2
28:00 259.7 F 21.3
30:00 269.7 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.6 F 29.9
2:00 271.6 F 29.0
3:00 272.4 F 30.0
4:00 272.2 F 29.4

Min 271.3 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 208.8 F 0.0
10:00 194.1 F 0.0
15:00 187.0 F 0.0
20:00 184.2 F 0.0
25:00 182.3 F 0.0
30:00 182.9 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:01
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:09:11
CYCLE COMPLETE

06/11/2024 09:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____