

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3629

Name: RITTER 1  
Sterilizer ID: V2178198  
08/09/2023 03:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.7 F 0.0  
2:00 178.7 F 3.3  
4:00 186.5 F 7.2  
6:00 217.4 F 11.9  
8:00 222.6 F 6.0  
10:00 234.2 F 10.6  
12:00 233.4 F 7.5  
14:00 242.6 F 12.8  
16:00 241.5 F 11.4  
18:00 253.4 F 17.4  
20:00 263.9 F 23.6  
22:00 271.9 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.8 F 29.3  
2:00 271.5 F 28.7  
3:00 271.7 F 29.1  
4:00 272.6 F 29.3

Min 271.0 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.5 F 0.0  
5:00 214.6 F 0.0  
10:00 229.8 F 0.0  
15:00 226.9 F 0.0  
20:00 225.2 F 0.0  
25:00 225.7 F 0.0  
30:00 220.7 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 22:11  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:01:22  
CYCLE COMPLETE

08/09/2023 04:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_