
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002650
8:10 10/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:09
STERILIZING 4:19
136.4 C 223kPa 4:19
136.2 C 222kPa 4:49
136.4 C 223kPa 5:19
136.4 C 223kPa 5:49
136.1 C 223kPa 6:19
136.4 C 219kPa 6:49
136.1 C 220kPa 7:19
136.3 C 222kPa 7:49
136.3 C 222kPa 7:49
VENTING 7:49
AIR DRYING 8:39
CYCLE COMPLETE 8:41

Digital Signature #
8D47A6D064FBC09B
