

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1623

Name: RITTER 1  
Sterilizer ID: V2178198  
02/11/2022 11:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 83.8 F 0.0  
2:00 97.3 F 0.0  
4:00 121.8 F 0.9  
6:00 154.2 F 2.9  
8:00 177.3 F 5.4  
10:00 195.3 F 8.7  
12:00 211.3 F 12.6  
14:00 219.6 F 5.2  
16:00 231.2 F 9.4  
18:00 237.4 F 10.5  
20:00 241.2 F 11.4  
22:00 237.9 F 9.4  
24:00 252.6 F 16.5  
26:00 264.7 F 23.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.3 F 28.7  
2:00 272.7 F 28.7  
3:00 272.1 F 28.3  
4:00 271.7 F 28.2

Min 271.6 F 28.0  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 231.7 F 0.0  
5:00 217.7 F 0.0  
10:00 216.9 F 0.0  
15:00 213.0 F 0.0  
20:00 210.3 F 0.0  
25:00 207.5 F 0.0  
30:00 205.0 F 0.0

Drying Complete

FILLING: 0:36  
HEATING: 27:40  
STERILIZING: 4:00  
VENTING: 2:45  
DRYING: 30:00  
TOTAL CYCLE: 01:05:01  
CYCLE COMPLETE

02/11/2022 12:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_