

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000564  
9:34 06/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:27  
PRESSURIZING 3:31  
STERILIZING 4:50  
136.0 C 220kPa 4:50  
135.8 C 228kPa 5:20  
136.4 C 232kPa 5:50  
136.4 C 232kPa 6:20  
136.4 C 232kPa 6:50  
136.3 C 230kPa 7:20  
136.4 C 232kPa 7:50  
136.4 C 232kPa 8:20  
136.4 C 232kPa 8:20  
VENTING 8:20  
AIR DRYING 9:15  
CYCLE COMPLETE 24:15  
-----

Digital Signature #  
03CEC4008C8C5A83  
-----