

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3386

Name: RITTER 1  
Sterilizer ID: V2178198  
06/14/2023 09:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.4 F 0.0  
2:00 182.2 F 0.0  
4:00 200.1 F 5.0  
6:00 217.7 F 10.1  
8:00 227.0 F 7.0  
10:00 237.9 F 11.5  
12:00 236.3 F 9.3  
14:00 237.9 F 9.2  
16:00 249.3 F 15.0  
18:00 261.4 F 21.8  
20:00 272.2 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.5 F 28.9  
2:00 271.3 F 28.3  
3:00 272.6 F 29.1  
4:00 271.9 F 28.5

Min 271.3 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.7 F 0.0  
5:00 210.7 F 0.0  
10:00 208.9 F 0.0  
15:00 208.8 F 0.0  
20:00 211.6 F 0.0  
25:00 210.9 F 0.0  
30:00 210.4 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 20:30  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 00:59:31  
CYCLE COMPLETE

06/14/2023 10:40 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_