
Midmark M11 - v1.0.0
Total Cycles: 4400

Name: RITTER 1
Sterilizer ID: V2178198
03/25/2024 08:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.0 F 0.0
2:00 71.6 F 0.2
4:00 80.8 F 1.0
6:00 101.3 F 2.8
8:00 129.3 F 5.2
10:00 152.9 F 8.0
12:00 179.0 F 11.1
14:00 216.7 F 4.2
16:00 216.7 F 4.2
18:00 225.8 F 7.1
20:00 232.5 F 10.3
22:00 236.2 F 10.1
24:00 239.2 F 10.1
26:00 238.7 F 9.3
28:00 246.0 F 12.8
30:00 255.7 F 18.2
32:00 265.3 F 24.0
34:00 271.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.2 F 28.4
2:00 272.7 F 28.7
3:00 272.0 F 28.9
4:00 271.3 F 28.3

Min 271.2 F 28.2
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.6 F 0.0
5:00 236.9 F 0.0
10:00 190.8 F 0.0
15:00 209.3 F 0.0
20:00 181.2 F 0.0
25:00 192.1 F 0.0
30:00 197.9 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 34:34
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:13:05
CYCLE COMPLETE

03/25/2024 09:57 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____