
Midmark M11 - v1.0.0
Total Cycles: 4606

Name: RITTER 1
Sterilizer ID: V2178198
06/04/2024 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.9 F 0.0
2:00 110.9 F 0.8
4:00 126.3 F 2.8
6:00 155.9 F 5.3
8:00 179.3 F 8.4
10:00 194.0 F 11.7
12:00 210.1 F 2.7
14:00 220.6 F 6.2
16:00 227.3 F 10.0
18:00 235.2 F 10.1
20:00 237.3 F 9.9
22:00 239.2 F 9.9
24:00 245.5 F 12.9
26:00 256.2 F 18.5
28:00 266.0 F 24.5
30:00 272.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 28.9
2:00 271.9 F 29.0
3:00 272.2 F 28.8
4:00 272.1 F 29.2

Min 271.3 F 28.5
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.7 F 0.0
5:00 232.2 F 0.0
10:00 193.9 F 0.0
15:00 185.3 F 0.0
20:00 185.3 F 0.0
25:00 185.0 F 0.0
30:00 183.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:31
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:09:13
CYCLE COMPLETE

06/04/2024 12:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____