
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005577
12:54 23/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:17
STERILIZING 4:15
136.4 C 223kPa 4:15
136.4 C 225kPa 4:45
136.4 C 225kPa 5:15
136.4 C 225kPa 5:45
136.2 C 224kPa 6:15
136.4 C 225kPa 6:45
136.4 C 223kPa 7:15
136.4 C 223kPa 7:45
136.4 C 223kPa 7:45
VENTING 7:45
AIR DRYING 8:37
CYCLE COMPLETE 18:30

Digital Signature #
B3A1C0ABC87F9ECB
