

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001938  
9:59 09/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:42  
PRESSURIZING 2:47  
STERILIZING 3:51  
136.0 C 220kPa 3:51  
136.2 C 223kPa 4:21  
136.3 C 224kPa 4:51  
136.4 C 224kPa 5:21  
136.4 C 224kPa 5:51  
136.4 C 223kPa 6:21  
136.0 C 221kPa 6:51  
136.2 C 222kPa 7:21  
136.2 C 222kPa 7:21  
VENTING 7:21  
AIR DRYING 8:09  
CYCLE COMPLETE 38:09

-----  
Digital Signature #  
7B78FC38969C9017  
-----