

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1906

Name: RITTER2  
Sterilizer ID: V2313228  
01/09/2023 01:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 132.8 F 0.0  
2:00 138.2 F 1.9  
4:00 165.3 F 5.0  
6:00 193.3 F 9.4  
8:00 215.3 F 7.0  
10:00 220.0 F 8.6  
12:00 231.1 F 11.2  
14:00 236.7 F 11.1  
16:00 235.8 F 9.5  
18:00 250.1 F 16.2  
20:00 262.2 F 23.3  
22:00 273.0 F 30.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.5  
1:00 272.2 F 30.4  
2:00 272.5 F 30.0  
3:00 271.8 F 29.5  
4:00 271.1 F 29.5

Min 271.1 F 29.3  
Max 272.7 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.8 F 0.0  
5:00 216.3 F 0.0  
10:00 243.8 F 0.0  
15:00 237.9 F 0.0  
20:00 227.1 F 0.0  
25:00 228.4 F 0.0  
30:00 219.6 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 22:25  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:00:34  
CYCLE COMPLETE

01/09/2023 02:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_