
Midmark M11 - v1.0.0
Total Cycles: 1774

Name: RITTER 1
Sterilizer ID: V2178198
03/31/2022 08:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.8 F 0.0
2:00 160.7 F 2.2
4:00 175.9 F 5.5
6:00 202.1 F 10.1
8:00 223.0 F 5.6
10:00 232.6 F 10.0
12:00 233.8 F 8.3
14:00 243.0 F 12.5
16:00 240.3 F 10.8
18:00 253.6 F 17.4
20:00 265.2 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.7 F 29.3
2:00 272.2 F 28.8
3:00 272.8 F 29.5
4:00 271.6 F 28.7

Min 271.0 F 28.3
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 230.0 F 0.0
10:00 236.2 F 0.0
15:00 217.9 F 0.0
20:00 209.4 F 0.0
25:00 218.0 F 0.0
30:00 210.0 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 21:52
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

03/31/2022 09:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____